


Masterpiece™ non-stick Bipolar Forceps

Please pass to Sterile Processing Department!



If necessary, use a soft cloth or brush.
Do not use a wire brush, scalpel, scratch sponge
or similar item.



Do not bend the forceps.



To protect the instrument please use
the storage trays.



The overall reconditioning process encompasses pre-cleaning, cleaning/disinfection, and sterilization.
Please refer to the instrument's instructions for use for more information and observe national guidelines
and regulations.